



Republic of the Philippines
Department of Health
MARGOSATUBIG REGIONAL HOSPITAL
Margosatubig, Zamboanga del Sur



(0981) 141-4419

mrh_gov@yahoo.com.ph/mrhgov@gmail.com

<http://mrh.doh.gov.ph>

Date: _____

Validity: 60 days

Minimum Technical Requirements for
BLUETOOTH SPEAKER

PRODUCT SPECIFICATIONS

- Battery (max power)
 - THD=30%
 - 78watts x 2
- Distortion
 - <1%
- Frequency
 - 20Hz – 20 kHz
- AUX sensitivity
 - 500 mV
- Guitar sensitivity
 - 50 mV
- Low voltage protection
 - 9.5V
- Charging current
 - 1.5A
- USB/MP3 Display format
 - WAV
 - WMA
 - MP3
 - FLAC
- Working time
 - 5 hours
- Bluetooth
 - 5.0
- Power requirements
 - 220-240Vac
 - Battery: 12Vdc 7.5amperes

Page | 1

PREPARED & REVIEWED BY

END USER: _____
DESIGNATION: _____

JAMES CARLTY D. LABADAN, ECE
CHAIRPERSON

KEVIN S. CATAGUE, ECE
VICE-CHAIRPERSON

JOEL L. LUNA
MEMBER

IVAN PAUL L. TANO, ECE
MEMBER

CRIS ABEL M. GUEVARRA, MPA
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS
Medical Center Chief II

Philhealth Accredited
REVISION: 7

FM-MCC-005

EFFECTIVITY DATE: APRIL 1, 2024



Republic of the Philippines
Department of Health
MARGOSATUBIG REGIONAL HOSPITAL
Margosatubig, Zamboanga del Sur



(0981) 141-4419

mrh_gov@yahoo.com.ph/mrhgov@gmail.com

<http://mrh.doh.gov.ph>

Date: _____

Validity: 60 days

Minimum Technical Requirements for
BLUETOOTH SPEAKER

STANDARD ACCESSORIES:

NOTE: accessories sets is per unit

- Wireless microphone

Page | 2

STANDARD REQUIREMENTS UPON SUBMISSION OF BIDDING DOCUMENTS

- Soft copy of user manual

REQUIREMENT/S IF AWARDED THE CONTRACT

- Unit must be **BRAND NEW**.
- Manual in English Language
 - 2 copy of Operation manual (Hard copy)
 - Service manual with soft copy
- Training Program
 - for end-user on equipment's operation with certificate
 - Technical training with certificate (When necessary)
- Three (3) years warranty on parts and service after commissioning and acceptance
- Certification on the availability of spare parts for the next 5 years
- Down time of machine will be added as extension of the warranty.
- With calibration certificate (When necessary)
- Supplier must respond to a service call before 72 hours and should be onsite not more than 7 days.
- **Section V. Special Conditions of the Contract**
 - Delivery Place: Margosatubig Regional Hospital, Zamboanga del Sur
- **Section VI Schedule of Requirements**

PREPARED & REVIEWED BY

END USER: _____
DESIGNATION: _____

JAMES CARLTY D. LABADAN, ECE
CHAIRPERSON

KEVIN S. CATAGUE, ECE
VICE-CHAIRPERSON

JOEL L. LUNA
MEMBER

IVAN PAUL L. TANO, ECE
MEMBER

CRIS ABEL M. GUEVARRA, MPA
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS
Medical Center Chief II

Philhealth Accredited
REVISION: 7

FM-MCC-005

EFFECTIVITY DATE: APRIL 1, 2024



Republic of the Philippines
Department of Health
MARGOSATUBIG REGIONAL HOSPITAL
Margosatubig, Zamboanga del Sur



(0981) 141-4419

mrh_gov@yahoo.com.ph/mrhgov@gmail.com

<http://mrh.doh.gov.ph>

Date: _____

Validity: 60 days

Minimum Technical Requirements for
BLUETOOTH SPEAKER

- Completion/Delivery Period: ____ calendar days upon receipt of the Notice to Proceed including the delivery, installation, testing and commissioning, Partial delivery allowed within the completion/delivery period.

Page | 3

PREPARED & REVIEWED BY

END USER: _____
DESIGNATION: _____

JAMES CARLTY D. LABADAN, ECE
CHAIRPERSON

KEVIN S. CATAGUE, ECE
VICE-CHAIRPERSON

JOEL L. LUNA
MEMBER

IVAN PAUL L. TANO, ECE
MEMBER

CRIS ABEL M. GUEVARRA, MPA
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS
Medical Center Chief II

Philhealth Accredited
REVISION: 7

FM-MCC-005

EFFECTIVITY DATE: APRIL 1, 2024