



Republic of the Philippines
Department of Health
MARGOSATUBIG REGIONAL HOSPITAL
Margosatubig, Zamboanga del Sur



(0981) 141-4419

mrh_gov@yahoo.com.ph/mrhgov@gmail.com

<http://mrh.doh.gov.ph>

Date: _____

Validity: 60 days

Minimum Technical Requirements for
HOSPITAL BED

PRODUCT SPECIFICATIONS

- Feature
 - Foldable wheel by crank
- Dimension
 - Length: 2100 to 2200mm
 - Width: 900 - 1000 mm
 - Variable Height: 400-600 mm
- Mattress
 - At least 120 mm thickness
- Backrest
 - Tilt angle: 0 to 65°
- Thigh section
 - Tilt angle: 0 to 35°
- Safe working load
 - 170 kgs
- Castor with brakes
- Collapsible side rail
- At least 2 hook IV pole
- Drainage bag rails
- Headboard and footboard
 - ABS material

Page | 1

STANDARD ACCESSORIES:

NOTE: accessories sets is per unit

- Overbed table
- IV pole
- Mattress

STANDARD REQUIREMENTS UPON SUBMISSION OF BIDDING DOCUMENTS

PREPARED & REVIEWED BY

END USER: _____
DESIGNATION: _____

JAMES CARLTY D. LABADAN, ECE
CHAIRPERSON

KEVIN S. CATAGUE, ECE
VICE-CHAIRPERSON

JOEL L. LUNA
MEMBER

IVAN PAUL L. TANO, ECE
MEMBER

CRIS ABEL M. GUEVARRA, MPA
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS
Medical Center Chief II

Philhealth Accredited
REVISION: 7

FM-MCC-005

EFFECTIVITY DATE: APRIL 1, 2024



Republic of the Philippines
Department of Health
MARGOSATUBIG REGIONAL HOSPITAL
Margosatubig, Zamboanga del Sur



(0981) 141-4419

mrh_gov@yahoo.com.ph/mrhgov@gmail.com

<http://mrh.doh.gov.ph>

Date: _____

Validity: 60 days

Minimum Technical Requirements for
HOSPITAL BED

➤ Soft copy of user manual
REQUIREMENT/S IF AWARDED THE CONTRACT

Page | 2

- Unit must be **BRAND NEW**.
- Manual in English Language
 - 2 copy of Operation manual (Hard copy)
 - Service manual with soft copy
- Training Program
 - for end-user on equipment's operation with certificate
 - Technical training with certificate (When necessary)
- Three (3) years warranty on parts and service after commissioning and acceptance
- Certification on the availability of spare parts for the next 5 years
- Down time of machine will be added as extension of the warranty.
- With calibration certificate (When necessary)
- Supplier must respond to a service call before 72 hours and should be onsite not more than 7 days.
- **Section V. Special Conditions of the Contract**
 - Delivery Place: Margosatubig Regional Hospital, Zamboanga del Sur
- **Section VI Schedule of Requirements**
 - Completion/Delivery Period: ____ calendar days upon receipt of the Notice to Proceed including the delivery, installation, testing and commissioning, Partial delivery allowed within the completion/delivery period.

PREPARED & REVIEWED BY

END USER: _____
DESIGNATION: _____

JAMES CARLTY D. LABADAN, ECE
CHAIRPERSON

KEVIN S. CATAGUE, ECE
VICE-CHAIRPERSON

JOEL L. LUNA
MEMBER

IVAN PAUL L. TANO, ECE
MEMBER

CRIS ABEL M. GUEVARRA, MPA
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS
Medical Center Chief II

Philhealth Accredited
REVISION: 7

FM-MCC-005

EFFECTIVITY DATE: APRIL 1, 2024