



## Republic of the Philippines Department of Health

## MARGOSATUBIG REGIONAL HOSPITAI



<u>ن</u> (	0981)	141-44	!1	S

mrh gov@yahoo.com.ph/mrhgov@gmail.com

l Sur			BAGONG		
1	Æ	http://mrh	doh	gov nh	

Date:	
Validity: 60 days	

### Minimum Technical Requirements for **PHOTOTHERAPY LIGHT**

#### **PRODUCT SPECIFICATIONS**

- Peak spectral irradiance
  - 55μW/cm2/nm at 45 cm
- Peak wavelength
  - o 451 nm
- **Emission spectrum** 
  - o 445 470 nm
- Effective surface area
  - At least 500x250 mm at recommended treatment distance
- Light intensity
  - Selectable (high or low)
- Lamp unit tilting
  - o Up to 90°
- Time monitoring
  - Therapy time monitoring
- Castor
  - Lockable 0
- Adjustable height
  - At least 150cm max height
- Power requirements
  - o 220 240Vac
- LED lifespan
  - o At least 60,000 hours
- Standards compliance
  - IEC 60601-1
  - IEC 60601-1-2
  - IEC 60601-2-50

PREPARED & REVIEWED BY

END USER: DESIGNATION:

JAMES CARLTY D. LABADAN, ECE

CHAIRPERSON

**KEVIN S. CATAGUE, ECE** VICE-CHAIRPERSON

JOEL L. LUNA

IVAN PAUL L. TANO, ECE MEMBER

CRIS ABEL M. GUEVARRA, MPA

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS

Medical Center Chief II

FM-MCC-005 **REVISION: 7** EFFECTIVITY DATE: APRIL 1, 2024

Page | 1





# Republic of the Philippines Department of Health

# MARGOSATUBIG REGIONAL HOSPITAI



(0981)	141-4419
--------	----------

mrh_	_gov@yahoo.com.ph/mrhgov@gmail
------	--------------------------------

	http://mrh.doh.g	iov.pl
₩	http://mrh.doh.g	10V. j

Date:	
Validity: 60 days	

#### Minimum Technical Requirements for **PHOTOTHERAPY LIGHT**

STANDARD ACCESSORIES:

NOTE: accessories sets is per unit

Bassinet with lockable wheels STANDARD REQUIREMENTS UPON SUBMISSION OF BIDDING DOCUMENTS

Page | 2

Soft copy of user manual

## REQUIREMENT/S IF AWARDED THE CONTRACT

- Unit must be BRAND NEW.
- Manual in English Language
  - 2 copy of Operation manual (Hard copy)
  - Service manual with soft copy
- **Training Program** 
  - for end-user on equipment's operation with certificate
  - Technical training with certificate (When necessary)
- > Three (3) years warranty on parts and service after commissioning and acceptance
- ➤ Certification on the availability of spare parts for the next 5 years
- > Down time of machine will be added as extension of the warranty.
- With calibration certificate (When necessary)
- > Supplier must respond to a service call before 72 hours and should be onsite not more than 7 days.
- **Section V. Special Conditions of the Contract** 
  - Delivery Place: Margosatubig Regional Hospital, Zamboanga del Sur
- **Section VI Schedule of Requirements**

PREPARED & REVIEWED BY

END USER: DESIGNATION:

JAMES CARLTY D. LABADAN, ECE CHAIRPERSON

**KEVIN S. CATAGUE, ECE** VICE-CHAIRPERSON

JOEL L. LUNA MEMBER

IVAN PAUL L. TANO, ECE MEMBER

CRIS ABEL M. GUEVARRA, MPA

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS Medical Center Chief II

> Philhealth Accredited **REVISION: 7**





# Republic of the Philippines Department of Health

## MARGOSATUBIG REGIONAL HOSPITAL Margosatubig, Zamboanga del Sur



Page | 3

_		
4.5		141-4419
C.	(0001)	111 1110
_	109011	141-4419

mrh\_gov@yahoo.com.ph/mrhgov@gmail.com

	http://mrh.doh.gov.ph
--	-----------------------

Date:	
Validity: 60 days	
	Minimum Technical Requirements for PHOTOTHERAPY LIGHT
•	Completion/Delivery Period: calendar days upon receipt of the Notice to Proceed including the delivery, installation, testing and commissioning, Partial delivery allowed within the completion/delivery period.

PREPARED & REVIEWED BY

END USER: \_\_\_\_\_\_
DESIGNATION: \_\_\_\_\_

JAMES CARLTY D. LABADAN, ECE CHAIRPERSON KEVIN S. CATAGUE, ECE VICE-CHAIRPERSON JOEL L. LUNA MEMBER IVAN PAUL L. TANO, ECE MEMBER CRIS ABEL M. GUEVARRA, MPA

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS

Medical Center Chief II