



Republic of the Philippines
Department of Health
MARGOSATUBIG REGIONAL HOSPITAL
Margosatubig, Zamboanga del Sur



(0981) 141-4419

mrh_gov@yahoo.com.ph/mrhgov@gmail.com

<http://mrh.doh.gov.ph>

Date: _____

Validity: 60 days

Minimum Technical Requirements for
Thermosealer

PRODUCT SPECIFICATIONS

- Built-in adjustable guide
 - 5 to 30mm
- Heat sealable materials
 - Pouches/reels
 - Tyvek pouch
 - Paper pouches
 - Gusseted pouches
 - Laminated AL pouches
 - PP pouches and ultra pouches
- Sealing
 - Rotary sealing
- Kind of sealing
 - Multiline
- Sealing width
 - 14 mm
- Distance sealing-medical product
 - $\geq 30\text{mm}$
- Sealing speed
 - 10 m/min
- Sealing temperature
 - 50 – 250 °C
- Sealing pressure
 - Manual control
- Interface
 - LCD
 - Keypad or equivalent technology
- Power supply

Page | 1

PREPARED & REVIEWED BY

END USER: _____
DESIGNATION: _____

JAMES CARLTY D. LABADAN, ECE
CHAIRPERSON

KEVIN S. CATAGUE, ECE
VICE-CHAIRPERSON

JOEL L. LUNA
MEMBER

IVAN PAUL L. TANO, ECE
MEMBER

CRIS ABEL M. GUEVARRA, MPA
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS
Medical Center Chief II

Philhealth Accredited
REVISION: 7

FM-MCC-005

EFFECTIVITY DATE: APRIL 1, 2024



Republic of the Philippines
Department of Health
MARGOSATUBIG REGIONAL HOSPITAL
Margosatubig, Zamboanga del Sur



(0981) 141-4419

mrh_gov@yahoo.com.ph/mrhgov@gmail.com

<http://mrh.doh.gov.ph>

Date: _____

Validity: 60 days

Minimum Technical Requirements for
Thermosealer

- 220 – 240 Vac

STANDARD ACCESSORIES:

Page | 2

NOTE: accessories sets is per unit

- Online UPS at least 1500 watts
- Roller tray or equivalent technology

STANDARD REQUIREMENTS UPON SUBMISSION OF BIDDING DOCUMENTS

- Soft copy of user manual

REQUIREMENT/S IF AWARDED THE CONTRACT

- Unit must be **BRAND NEW**.
- Manual in English Language
 - 2 copy of Operation manual (Hard copy)
 - Service manual with soft copy
- Training Program
 - for end-user on equipment's operation with certificate
 - Technical training with certificate (When necessary)
- Three (3) years warranty on parts and service after commissioning and acceptance
- Installation and commissioning
- Trained Technical personnel to conduct preventive and corrective action of the machine.
- Certification on the availability of spare parts for the next 5 years

PREPARED & REVIEWED BY

END USER: _____
DESIGNATION: _____

JAMES CARLTY D. LABADAN, ECE
CHAIRPERSON

KEVIN S. CATAGUE, ECE
VICE-CHAIRPERSON

JOEL L. LUNA
MEMBER

IVAN PAUL L. TANO, ECE
MEMBER

CRIS ABEL M. GUEVARRA, MPA
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS
Medical Center Chief II

Philhealth Accredited
REVISION: 7

FM-MCC-005

EFFECTIVITY DATE: APRIL 1, 2024



Republic of the Philippines
Department of Health
MARGOSATUBIG REGIONAL HOSPITAL
Margosatubig, Zamboanga del Sur



(0981) 141-4419

mrh_gov@yahoo.com.ph/mrhgov@gmail.com

<http://mrh.doh.gov.ph>

Date: _____

Validity: 60 days

Minimum Technical Requirements for
Thermosealer

- Down time of machine will be added as extension of the warranty.
- With calibration certificate (When necessary)
- Supplier must respond to a service call before 72 hours and should be onsite not more than 7 days.
- **Section V. Special Conditions of the Contract**
 - Delivery Place: Margosatubig Regional Hospital, Zamboanga del Sur
- **Section VI Schedule of Requirements**
 - Completion/Delivery Period: ____ calendar days upon receipt of the Notice to Proceed including the delivery, installation, testing and commissioning, Partial delivery allowed within the completion/delivery period.

Page | 3

PREPARED & REVIEWED BY

END USER: _____
DESIGNATION: _____

JAMES CARLTY D. LABADAN, ECE
CHAIRPERSON

KEVIN S. CATAGUE, ECE
VICE-CHAIRPERSON

JOEL L. LUNA
MEMBER

IVAN PAUL L. TANO, ECE
MEMBER

CRIS ABEL M. GUEVARRA, MPA
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS
Medical Center Chief II

Philhealth Accredited
REVISION: 7

FM-MCC-005

EFFECTIVITY DATE: APRIL 1, 2024