



Republic of the Philippines  
Department of Health  
**MARGOSATUBIG REGIONAL HOSPITAL**  
Margosatubig, Zamboanga del Sur



(0981) 141-4419

[mrh\\_gov@yahoo.com.ph/mrhgov@gmail.com](mailto:mrh_gov@yahoo.com.ph/mrhgov@gmail.com)

<http://mrh.doh.gov.ph>

Date: \_\_\_\_\_

Validity: 60 days

Minimum Technical Requirements for  
**TRANSFER STRETCHER**

**PRODUCT SPECIFICATIONS**

- Dimension
  - Bed length
    - Not longer than 1950mm
  - Width with side rails
    - Not wider than 670mm
  - Height range
    - 510 – 850 mm
- Articulates
  - Back section
    - 0 to 70°
- Casters
  - Not smaller than 150 mm with central locking
  - Center wheel: not larger 100 mm diameter
- Maximum patient weight
  - Not lower than 135 kg
- Mattress
  - Polyester cover
  - Urethane foam
- Material
  - Frame
    - Steel
  - Mattress base
    - Polypropylene
  - Transfer board
    - Polypropylene
- Oxygen holder capacity
  - 500L (100 to 115 mm diameter)

Page | 1

PREPARED & REVIEWED BY

END USER: \_\_\_\_\_  
DESIGNATION: \_\_\_\_\_

JAMES CARLTY D. LABADAN, ECE  
CHAIRPERSON

KEVIN S. CATAGUE, ECE  
VICE-CHAIRPERSON

JOEL L. LUNA  
MEMBER

IVAN PAUL L. TANO, ECE  
MEMBER

CRIS ABEL M. GUEVARRA, MPA  
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS  
Medical Center Chief II

*Philhealth Accredited*

FM-MCC-005

REVISION: 7

EFFECTIVITY DATE: APRIL 1, 2024



Republic of the Philippines  
Department of Health  
**MARGOSATUBIG REGIONAL HOSPITAL**  
Margosatubig, Zamboanga del Sur



(0981) 141-4419

[mrh\\_gov@yahoo.com.ph/mrhgov@gmail.com](mailto:mrh_gov@yahoo.com.ph/mrhgov@gmail.com)

<http://mrh.doh.gov.ph>

Date: \_\_\_\_\_

Validity: 60 days

Minimum Technical Requirements for  
**TRANSFER STRETCHER**

**STANDARD ACCESSORIES:**

**NOTE: accessories sets is per unit**

- IV pole
- Oxygen cylinder holder
- 4 pcs hook

Page | 2

**STANDARD REQUIREMENTS UPON SUBMISSION OF BIDDING DOCUMENTS**

- Soft copy of user manual

**REQUIREMENT/S IF AWARDED THE CONTRACT**

- Unit must be **BRAND NEW**.
- Manual in English Language
  - 2 copy of Operation manual (Hard copy)
  - Service manual with soft copy
- Training Program
  - for end-user on equipment's operation with certificate
  - Technical training with certificate (When necessary)
- Three (3) years warranty on parts and service after commissioning and acceptance
- Certification on the availability of spare parts for the next 5 years
- Down time of machine will be added as extension of the warranty.
- With calibration certificate (When necessary)
- Supplier must respond to a service call before 72 hours and should be onsite not more than 7 days.
- **Section V. Special Conditions of the Contract**
  - Delivery Place: Margosatubig Regional Hospital, Zamboanga del Sur
- **Section VI Schedule of Requirements**

PREPARED & REVIEWED BY

END USER: \_\_\_\_\_  
DESIGNATION: \_\_\_\_\_

JAMES CARLTY D. LABADAN, ECE  
CHAIRPERSON

KEVIN S. CATAGUE, ECE  
VICE-CHAIRPERSON

JOEL L. LUNA  
MEMBER

IVAN PAUL L. TANO, ECE  
MEMBER

CRIS ABEL M. GUEVARRA, MPA  
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS  
Medical Center Chief II

*Philhealth Accredited*

FM-MCC-005

REVISION: 7

EFFECTIVITY DATE: APRIL 1, 2024



Republic of the Philippines  
Department of Health  
**MARGOSATUBIG REGIONAL HOSPITAL**  
Margosatubig, Zamboanga del Sur



(0981) 141-4419

[mrh\\_gov@yahoo.com.ph/mrhgov@gmail.com](mailto:mrh_gov@yahoo.com.ph/mrhgov@gmail.com)

<http://mrh.doh.gov.ph>

Date: \_\_\_\_\_

Validity: 60 days

Minimum Technical Requirements for  
**TRANSFER STRETCHER**

- Completion/Delivery Period: \_\_\_\_ calendar days upon receipt of the Notice to Proceed including the delivery, installation, testing and commissioning, Partial delivery allowed within the completion/delivery period.

Page | 3

PREPARED & REVIEWED BY

END USER: \_\_\_\_\_  
DESIGNATION: \_\_\_\_\_

**JAMES CARLTY D. LABADAN, ECE**  
CHAIRPERSON

**KEVIN S. CATAGUE, ECE**  
VICE-CHAIRPERSON

**JOEL L. LUNA**  
MEMBER

**IVAN PAUL L. TANO, ECE**  
MEMBER

**CRIS ABEL M. GUEVARRA, MPA**  
MEMBER

APPROVED BY:

**RICHARD B. SISON, MD, MHA, FPCHA, FPSMS**  
Medical Center Chief II

*Philhealth Accredited*  
REVISION: 7

FM-MCC-005

EFFECTIVITY DATE: APRIL 1, 2024