



Republic of the Philippines  
Department of Health  
**MARGOSATUBIG REGIONAL HOSPITAL**  
Margosatubig, Zamboanga del Sur



(0981) 141-4419

[mrh\\_gov@yahoo.com.ph/mrhgov@gmail.com](mailto:mrh_gov@yahoo.com.ph/mrhgov@gmail.com)

<http://mrh.doh.gov.ph>

Date: \_\_\_\_\_

Validity: 60 days

Minimum Technical Requirements for  
**VEIN FINDER**

**PRODUCT SPECIFICATIONS**

- Display
  - LCD display
- Light source
  - Infrared
- Display mode
  - Projection
- Infrared wavelength
  - 850 nm
- Imaging depth of field
  - At least 30mm
- Infrared radiation energy
  - $\leq 0.6 \text{ mW/m}^2$
- Power supply
  - Rechargeable battery

Page | 1

**STANDARD ACCESSORIES:**

**NOTE: accessories sets is per unit**

**STANDARD REQUIREMENTS UPON SUBMISSION OF BIDDING DOCUMENTS**

- Soft copy of user manual

**REQUIREMENT/S IF AWARDED THE CONTRACT**

- Unit must be **BRAND NEW**.
- Manual in English Language

PREPARED & REVIEWED BY

END USER: \_\_\_\_\_  
DESIGNATION: \_\_\_\_\_

JAMES CARLTY D. LABADAN, ECE  
CHAIRPERSON

KEVIN S. CATAGUE, ECE  
VICE-CHAIRPERSON

JOEL L. LUNA  
MEMBER

IVAN PAUL L. TANO, ECE  
MEMBER

CRIS ABEL M. GUEVARRA, MPA  
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS  
Medical Center Chief II

*Philhealth Accredited*  
REVISION: 7

FM-MCC-005

EFFECTIVITY DATE: APRIL 1, 2024



Republic of the Philippines  
Department of Health  
**MARGOSATUBIG REGIONAL HOSPITAL**  
Margosatubig, Zamboanga del Sur



(0981) 141-4419

[mrh\\_gov@yahoo.com.ph/mrhgov@gmail.com](mailto:mrh_gov@yahoo.com.ph/mrhgov@gmail.com)

<http://mrh.doh.gov.ph>

Date: \_\_\_\_\_

Validity: 60 days

Minimum Technical Requirements for  
**VEIN FINDER**

- 2 copy of Operation manual (Hard copy)
- Service manual with soft copy
- Training Program
  - for end-user on equipment's operation with certificate
  - Technical training with certificate (When necessary)
- Three (3) years warranty on parts and service after commissioning and acceptance
- Certification on the availability of spare parts for the next 5 years
- Down time of machine will be added as extension of the warranty.
- With calibration certificate (When necessary)
- Supplier must respond to a service call before 72 hours and should be onsite not more than 7 days.
- **Section V. Special Conditions of the Contract**
  - Delivery Place: Margosatubig Regional Hospital, Zamboanga del Sur
- **Section VI Schedule of Requirements**
  - Completion/Delivery Period: \_\_\_\_\_ calendar days upon receipt of the Notice to Proceed including the delivery, installation, testing and commissioning, Partial delivery allowed within the completion/delivery period.

Page | 2

PREPARED & REVIEWED BY

END USER: \_\_\_\_\_  
DESIGNATION: \_\_\_\_\_

JAMES CARLTY D. LABADAN, ECE  
CHAIRPERSON

KEVIN S. CATAGUE, ECE  
VICE-CHAIRPERSON

JOEL L. LUNA  
MEMBER

IVAN PAUL L. TANO, ECE  
MEMBER

CRIS ABEL M. GUEVARRA, MPA  
MEMBER

APPROVED BY:

RICHARD B. SISON, MD, MHA, FPCHA, FPSMS  
Medical Center Chief II